AsureQuality Kaitiaki Kai

MICROBIOLOGICAL MEDIA
QUALITY PERFORMANCE REPORT # 26087

Test Media: Rapid Yeast and Mould Count Plate

Brand: Neogen Petrifilm™ **Batch No. / Expiry Date:** 3497T7/2027-11-05

Reference Media: Yeast Dextrose Chloramphenicol (YDC) Agar

Brand: Millipore **Batch No. / Expiry Date:** VM1132900 / Apr 2030

Date Tested: 02 Jul 2026 **Project No:** 26-185925-1A

BATCH PERFORMANCE CERTIFICATE

Qualitative Analysis

Test Organism	Observed Growth
Staphylococcus aureus NZRM 87	No Growth
Escherichia coli NZRM 916	No Growth
Saccharomyces cerevisiae NZRM 2241	Positive
Pseudomonas aeruginosa NZRM 918	No Growth
Aspergillus brasiliensis NZRM 2578	Positive
Blank	No Growth

Positive: Typical growth as described by method.

Negative: Growth not described as typical.

No Growth: No visible colonies observed.

Method: Direct inoculation i.e. 1ml of a diluted overnight culture was used to inoculate the test media.

Blank: 5 plates inoculated with 1ml sterile diluent.

Incubation Conditions: Aerobic 25°C/2 days

Quantitative Analysis

-	Test Organism	
-	*Saccharomyces cerevisiae* NZRM 2241	-
Ave. No. Colonies Petrifilm	108.2	-
Ave. No. Colonies YDC	107.4	-
% Difference	0.7%	-

Quantitative analysis compares the recovery rate of the test media to a reference media.

Calculation: The percentage difference =

$$\frac{[\text{Ave. No. of Test Media Colonies} \times 100]}{\text{Ave. No. of Reference Media Colonies}} - 100$$

Method: In House

Incubation Conditions: Aerobic 25°C/2 days

Quantitative Interpretation: The media performance is considered acceptable if the percentage difference between the test and non-selective reference media does not exceed ±10%.

Qualitative Interpretation: The media performance is considered acceptable if both positive growth of the target organism(s) and negative or no growth of non-target organism(s) is expressed.

The above criteria are testing laboratory In-house limits, which have been endorsed by Neogen as a report interpretation guideline.

Results apply to samples of the Batch as received by the laboratory.

Authorised by:

Michele An
Scientific Analyst

Report Date: 07 July 2026

This report must not be altered, or reproduced except in full.

AsureQuality Limited
131 Boundary Road, Blockhouse Bay, Auckland, New Zealand
Telephone (09) 626 8000; Fax (09) 626 8282